



GOVERNMENT OF WEST BENGAL
IRRIGATION AND WATERWAYS DEPARTMENT
NATIONAL HYDROLOGY PROJECT

NAME OF WORK:-“Procurement of Water Quality Laboratory Instruments including Comprehensive Annual Maintenance Contract for 3 Years (After 1 Year Warranty) under National Hydrology Project”.

Bid No.: WBIW/NHP/IFB-02/2021-22

(Circulated vide this Office Memo no 225/NHP-02/02. Dated 06.07.2021)

CORRIGENDUM NOTICE- 2


Due to unavoidable circumstances, the following changes have been made: -

Turbidity Meter (Page-31 of Bid document):-

Specification Parameters	As per published Bid document	Now to be read as
Measurement Range:	0 – 4000NTU [Range: 0 – 40 (Ratio off) & 0 – 4000 NTU (Ratio on)]	0 - 1100 NTU
Certification:	CE, KC, RCM	CE, ISO
Communication:	USB	USB/ RS 232
Display:	17.8mm color touch screen	Touch Screen/ Key Pad Operation
Interface:	USB-A ports for USB flash drive	USB-A ports for USB flash drive/ RS 232
Light Source:	Tungsten filament lamp	Infra Red Light source
Reading Modes:	Single, continuous, Rapidly Settling Turbidity™, signal averaging on or off, ratio on or off	Single, continuous
Response Time:	6.8 Seconds in ratio off mode/14 seconds in ratio on mode taking 10 measurements at a time	6.8 Seconds or better
Stabilization Time:	Ratio on: 30 minutes after start-up/	30 minutes after start-up

All other Terms & Conditions of the Bid will remain unchanged.

Date:-29/07/2021


Executive Engineer
Burdwan Investigation & Planning Division
I & W Dte., Govt. of West Bengal